



## SPECTRORADIOMETER CALIBRATION CERTIFICATE

<b>Measuring Object</b>	Array Type Spectroradiometer
<b>Manufacturer</b>	LMT Lichtmesstechnik GmbH Berlin
<b>Type</b>	LMT GO-DS SM 8107
<b>Serial No.</b>	07B316-81
<b>Customer</b>	Belintegra Joint-Stock Co.. 63 M. Lynkov Str. 220124 Minsk (Republic of Belarus)
<b>Order No.</b>	11B214
<b>Date of measurement</b>	November 05, 2013
<b>Kind of Measurement</b>	Calibration of the spectroradiometer as a system component of the goniophotometer with mirror arrangement against a calibrated incandescent halogen lamp
<b>Test Instruments</b>	LMT GO-DS SM 8107 Spectroradiometer SN 07B316-81 LMT GO-DS 2000 Goniophotometer SN 05B245 Incandescent Halogen Lamp OSRAM FEL/120 1000W SN 07B316-85 with LMT calibration certificate for colour temperature at rated operation current Power Supply Chroma 6415, SN 641500000509 Wattmeter Yokogawa WT 3000, SN 91MB10881
<b>Measurement Conditions</b>	The measurement was performed in draught-free conditions, ambient temperature $25^{\circ} \pm 1^{\circ} \text{C}$ .
<b>Calibration Procedure</b>	<b>Using LMT proprietary Calibration Software "LIMES 2000-Calibration"</b> . The incandescent halogen lamp was mounted on the GO-DS in base-up orientation with the cooling ribs of its measuring socket aligned at C-Plane $C=0^{\circ}$ (Offset $C=339,3^{\circ}$ ), $\gamma$ -Angle $\gamma=90^{\circ}$ and the lamp-filament holders aligned at C-Plane $C=180^{\circ}$ . By connection to the power supply the lamp was operated at a constant current of 8,158 A. As per lamp certificate under these conditions the lamp spectral distribution is equal to a correlated color temperature of 3100 K. The software of the LIMES 2000 generated automatically a calibration file based on measured raw data of the lamp spectrum as well as a theoretical 3100 K Planck spectral distribution.
<b>Measurement results</b>	$3100 \pm 30\text{K}$
<b>Remarks</b>	For subsequent measurements of light sources intended for general lighting applications including solid-state light sources the measurement uncertainty is estimated at up to $\pm 50\text{K}$ . It is recommended to validate the calibration files by means of the procedure described above at intervals of at least 30 days. A deviation of $\pm 30\text{K}$ is acceptable without need of re-calibration

Baryssau, November 05, 2013

**LMT LICHTMESSTECHNIK GMBH BERLIN**

Dipl.-Ing. Zulkifli  
System Engineer

